

Method Parameters

Analysis Lines

Label	El	Wavelen.	Type	I/S	Bkg Mode	PPP	OBCL	OBCR
Ag 328.068	Ag	328.068	Analyte	-	Fitted	2		

Conditions Sets (All lines share a single condition set)

Pwr(kW)	PlasFlow(L/min)	AuxFlow(L/min)	NebFlow(L/min)	Replicate Time(s)	Stab Time(s)	Multi Frame
1.20	15.0	1.50	1.00	5.000	15	On

Sample Introduction

Sample Uptake(s)	Rinse Time(s)	Pump Rate(rpm)	Fast Pump
30	10	15	On

General Settings

Replicates
3

Calibration Settings

Calib. Mode	No. Standards	Corr Coeff Limit	Reslopes
Quantitative	1	0.995000	Off

Standard Concentration

El	Units	Silver
Ag	ppm	30.00000

Calibration Parameters

Label	Thru. Blk	W/Fit	Curve Type	Max Error	Min Conc	Max Conc	E/Curve %	U/Curve %
Ag 328.068	Off	On	Linear	5.0	0.00000	33.00000	25	400

Units

Unit Name	Multiplier	Divider
mg/L	1	1E6
ppm	1	1E6

Columns

Column
Ag 328.068

Display Column Flags

Column	o	QC	S	e	uncal	I	n	a	u	v	x	b	p	m	d
Ag 328.068	x	x	x	x	x	x	x	x	x	x	x	x	x	x	x


Blank (Blk)

12/9/2015, 11:37:58 AM

Tube 1


Label	Sol'n Conc.	Units	SD(Int)	%RSD(Int)	Int. (c/s)
Ag 328.068	0.000000	ppm	16.745	13.5	123.730

Ionosil-CS-10-151116. All Data Report 12/10/2015, 8:20:55 AM



Label	Sol'n Conc.	Units	SD(Int)	%RSD(Int)	Int. (c/s)
Silver (Std)	30.0000	ppm	3300.442	0.3	963303

12/9/2015, 11:59:45 AM Tube 2




Label	Flags	Int. (c/s)	Std Conc.	Calc Conc.	Error	%Error
Blank		123.730	0.000000	0.000000	-	-
Silver		963303	30.0000	30.0000	0.000000	0.0

Ag 328.068 Calibration (ppm) 12/9/2015, 11:59:45 AM Correlation Coefficient: 1.000000

Curve Type: Linear Equation: $y = 32106.0 x + 123.7$


Label	Sol'n Conc.	Units	SD	%RSD	Int. (c/s)	Calc Conc.	DF
Ag 328.068	10.5513	ppm	0.028913	0.3	338883	10.5513 ppm	1.00000

Total Silver 1: (Samp) 12/9/2015, 11:57:23 AM Tube 3
 Weight: 1 Volume: 1 Dilution: 1




Label	Sol'n Conc.	Units	SD	%RSD	Int. (c/s)	Calc Conc.	DF
Ag 328.068	10.5927	ppm	0.016993	0.2	340212	10.5927 ppm	1.00000

Total Silver 2: (Samp) 12/9/2015, 11:55:12 AM Tube 4
 Weight: 1 Volume: 1 Dilution: 1



Ionosil-CS-10-151116. All Data Report 12/10/2015, 8:20:55 AM

Ionic Silver 1: (Samp)	12/9/2015, 11:49:43 AM	Tube 5					
Weight: 1	Volume: 1	Dilution: 1					
Label	Sol'n Conc.	Units	SD	%RSD	Int. (c/s)	Calc Conc.	DF
Ag 328.068	7.19739	ppm	0.017040	0.2	231203	7.19739 ppm	1.00000



Ionic Silver 2: (Samp)	12/9/2015, 11:51:32 AM	Tube 6					
Weight: 1	Volume: 1	Dilution: 1					
Label	Sol'n Conc.	Units	SD	%RSD	Int. (c/s)	Calc Conc.	DF
Ag 328.068	7.23772	ppm	0.031203	0.4	232498	7.23772 ppm	1.00000

